



ORDERING INFORMATION

9.5F Attachable	MRDP95ADNCA	W/ Micro Introducer	Silicone Filled Suture Holes W/ Micro Introducer	5/BOX
		MIDP95ADNCA	MIDP95ADSCA	

SETS WITH MICRO-STICK® CONTAIN:

- (1) Catheter
- (1) CT Implantable Port
- (2) Catheter Locks
- (1) Scalpel
- (1) Vein Pick
- (1) .035" x 70CM "J" Marked Guidewire with Advancer
- (1) 10cc Syringe
- (1) Valved Peelable Introducer
- (1) Tunneler
- (2) 22 Ga. Huber Needles (1-Straight, 1-Right Angle)
- (1) Blunt Tip Needle
- (1) Patient Information Pack
- (1) Patient Chart Sticker

MICRO-STICK®

- (1) 5FR Coaxial Dilator
- (1) .018" x 45 cm Guidewire with Radiopaque Tip
- (1) 21 Ga. Introducer Needle with Echogenic Tip

SETS WITHOUT MICRO-STICK® CONTAIN:

- (1) Catheter
- (1) CT Implantable Port
- (2) Catheter Locks
- (1) Scalpel
- (1) Vein Pick
- (1) Introducer Needle
- (1) .035" X 70CM "J" Marked Guidewire with Advancer
- (1) 10cc Syringe
- (1) Valved Peelable Introducer
- (1) Tunneler
- (2) 22ga Huber Needles (1-Straight, 1-Right Angle)
- (1) Blunt Tip Needle
- (1) Patient Information Pack
- (1) Patient Chart Sticker

Refer to the port kit IFU for procedural methods, contraindications and warnings.

INDICATED CT FLOW RATES

Non-Coring Needle Gauge Size	19 Ga.	20 Ga.	22 Ga.
Maximum Flow Rate*	5 ml/s	5 ml/s	2 ml/s

IMPORTANT RISK INFORMATION

Indications for Use: The Dignity® Dual Port is a power injectable implantable infusion port that is indicated for patient therapies requiring repeated access to the vascular system. The Dignity® Dual Port can be used for infusion of medications, I.V. fluids, parenteral nutrition solutions, blood products, and for the withdrawal of blood samples. When used with a power injectable needle, the Dignity® Dual Port is indicated for power injection of contrast media. For power injection of contrast media, the maximum recommended infusion rate is 5 ml/s with a 19 or 20 gauge non-coring power injectable needle. The maximum recommended infusion rate is 2 ml/s with a 22 gauge non-coring power injectable needle.

Contraindications: This device is contraindicated for catheter insertion in the subclavian vein medial to the border of the first rib, an area which is associated with higher rates of pinch-off. The device is also contraindicated: When the presence of device related infection, bacteremia, or septicemia is known or suspected. When the patient's body size is insufficient for the size of the implanted device. When the patient is known or is suspected to be allergic to materials contained in the device. If severe chronic obstructive lung disease exists. If the prospective insertion site has been previously irradiated. If the prospective placement site has previously suffered episodes of venous thrombosis or vascular surgical procedures. If local tissue factors will prevent proper device stabilization and/or access.

Refer to Instructions for Use provided with the product for complete instructions, warnings, and contraindications. Observe all Instructions for Use prior to using products. Failure to do so may result in patient complications.

Dignity®

Dual



Multiple therapies, same trusted port.

Dignity[®]

Dual


Multiple therapies, same trusted port.

DUAL LUMENS

Allow for multiple infusion therapies.

TAPERED TAIL

Aids in the tissue dissection process, creating a tighter port pocket which can help prevent flipping.



SMOOTH CONTOURS

RAISED RIM

CONCAVE SIDES



REDUCED WIDTH

Minimizes incision length during implantation and increases the septum-to-body ratio to help promote ease of access.



INLINE DESIGN

Minimizes incision length during implantation.



LOCKING MECHANISM.



RADIOPAQUE INK

"CT" printing can be visualized under x-ray to provide confirmation of power injectability.

