

# SUMMARY OF SAFETY AND CLINICAL PERFORMANCE

SSCP-007

## Split Stream® Catheter Sets Product Family

### IMPORTANT INFORMATION

This Summary of Safety and Clinical Performance (SSCP) is intended to provide public access to an updated summary of the main aspects of the safety and clinical performance of the device.

This SSCP is not intended to replace the Instructions for Use as the main document to ensure the safe use of the device, nor is it intended to provide diagnostic or therapeutic suggestions to intended users or patients.

Applicable Documents	
Document Type	Document Title / Number
DHF	02020, 03006, 03006-A2
'MDR Documentation' File Number	MDR-007

Revision History					
Revision	Date	CR#	Author	Description of Changes	Validated
1	04OCT2021	26535	RS	Implementation of SSCP	<input type="checkbox"/> Yes, this version was validated by the Notified Body in the following language: English <input type="checkbox"/> No, this version was not validated by the Notified Body as this is a Class IIa or IIb implantable device
2	28JUN2022	27030	RS	Scheduled Update	<input checked="" type="checkbox"/> Yes, this version was

Revision History					
Revision	Date	CR#	Author	Description of Changes	Validated
					validated by the Notified Body in the following language: English <input type="checkbox"/> No, this version was not validated by the Notified Body as this is a Class IIa or IIb implantable device
3	26JUN2023	28249	GM	Periodic Update; Updated in Accordance with CER-007, Revision D	<input type="checkbox"/> Yes, this version was validated by the Notified Body in the following language: English <input type="checkbox"/> No, this version was not validated by the Notified Body as this is a Class IIa or IIb implantable device
4	21JUN2024	29455	GM	Periodic Update; Updated in Accordance with CER-007, Revision E	<input type="checkbox"/> Yes, this version was validated by the Notified Body in the following language: English <input type="checkbox"/> No, this version was not validated by the Notified Body as this is a Class IIa or IIb implantable device

<b>Revision History</b>					
<b>Revision</b>	<b>Date</b>	<b>CR#</b>	<b>Author</b>	<b>Description of Changes</b>	<b>Validated</b>
<b>5</b>	<b>05SEP2025</b>	<b>25-0166</b>	<b>GM</b>	<b>Periodic Update; Updated in Accordance with CER-007, Revision F</b>	<input type="checkbox"/> Yes, this version was validated by the Notified Body in the following language: English <input type="checkbox"/> No, this version was not validated by the Notified Body as this is a Class IIa or IIb implantable device

---

**USERS / HEALTHCARE PROFESSIONALS**

---

The following information is intended for users/healthcare professionals. Following this information there is a summary intended for patients.

**1. Device identification and general information**

Device trade name(s)	Split Stream®
Manufacturer name and address	Medical Components, Inc. 1499 Delp Drive Harleysville, PA 19438 USA
Manufacturer single registration number (SRN)	US-MF-000008230
Basic UDI-DI	00884908249NH
Medical device nomenclature description / text	F900202 – Permanent Hemodialysis Catheter and Kits
Class of device	III
Date first CE certificate was issued for this device	2003
Authorized representative name and SRN	European Regulatory Expert Medical Product Service GmbH (MPS) Borngasse 20 35619 Braunfels, Germany SRN: DE-AR-000005009
Notified Body name and single identification number	BSI Netherlands NB2797

The devices in scope of this document are all long-term hemodialysis catheter sets. The device part numbers are organized into variant categories. These devices are distributed as procedure trays, in various configurations inclusive of accessories and adjunctive devices (see section “Accessories intended for use in combination with the Device”).

Variant Devices:

Variant Description	Part Number	Explanation of Multiple Part Numbers
14F x 24cm Split Stream	10298-824 10028-824	No significant clinical, biological, or technical difference (only difference is pre-loaded stylet)
14F x 28cm Split Stream	10298-828 10028-828	No significant clinical, biological, or technical difference (only difference is pre-loaded stylet)
14F x 32cm Split Stream	10298-832 10028-832	No significant clinical, biological, or technical difference (only difference is pre-loaded stylet)
14F x 36cm Split Stream	10298-836 10028-836	No significant clinical, biological, or technical difference (only difference is pre-loaded stylet)

Variant Description	Part Number	Explanation of Multiple Part Numbers
14F x 40cm Split Stream	10298-840 10028-840	No significant clinical, biological, or technical difference (only difference is pre-loaded stylet)
16F x 24cm Split Stream	10062-824	N/A
16F x 28cm Split Stream	10062-828	N/A
16F x 32cm Split Stream	10062-832	N/A
16F x 36cm Split Stream	10062-836	N/A
16F x 40cm Split Stream	10062-840	N/A

Procedure Trays:

Catalog Code	Part Number	Description
S24SE.	10298-824	14F x 24cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S28SE.	10298-828	14F x 28cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S32SE.	10298-832	14F x 32cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S36SE.	10298-836	14F x 36cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S40SE.	10298-840	14F x 40cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
SST24SE.	10298-824	14F x 24cm Split Stream® Catheter w/ Stylet Set (Cuff 19cm From Tip)
SST28SE.	10298-828	14F x 28cm Split Stream® Catheter w/ Stylet Set (Cuff 23cm From Tip)
SST32SE.	10298-832	14F x 32cm Split Stream® Catheter w/ Stylet Set (Cuff 27cm From Tip)
SST36SE.	10298-836	14F x 36cm Split Stream® Catheter w/ Stylet Set (Cuff 31cm From Tip)
SST40SE.	10298-840	14F x 40cm Split Stream® Catheter w/ Stylet Set (Cuff 35cm From Tip)
SST24E.	10028-824	14F x 24cm Split Stream® Catheter Set (Cuff 19cm From Tip)
SST28E.	10028-828	14F x 28cm Split Stream® Catheter Set (Cuff 23cm From Tip)
SST32E.	10028-832	14F x 32cm Split Stream® Catheter Set (Cuff 27cm From Tip)
SST36E.	10028-836	14F x 36cm Split Stream® Catheter Set (Cuff 31cm From Tip)
SST40E.	10028-840	14F x 40cm Split Stream® Catheter Set (Cuff 35cm From Tip)
SST2416E.	10062-824	16F x 24cm Split Stream® Catheter Set (Cuff 19cm From Tip)
SST2816E.	10062-828	16F x 28cm Split Stream® Catheter Set (Cuff 23cm From Tip)
SST3216E.	10062-832	16F x 32cm Split Stream® Catheter Set (Cuff 27cm From Tip)
SST3616E.	10062-836	16F x 36cm Split Stream® Catheter Set (Cuff 31cm From Tip)
SST4016E.	10062-840	16F x 40cm Split Stream® Catheter Set (Cuff 35cm From Tip)

Configurations of Procedure Trays:

Configuration Type	Kit Components
24cm and 28cm length Sets	<ul style="list-style-type: none"> <li>(1) Catheter</li> <li>(1) 1.3mm OD x 1.0mm ID x 70mm (18GA) INTRODUCER NEEDLE</li> <li>(1) 0.97mm x 70cm (.038) GUIDEWIRE J (R 3mm) TIP</li> <li>(1) Advancer</li> <li>(1) 3.3mm OD x 15cm 12° BENT Raulerson Ring Handled Y Adaptor Tunneler</li> <li>(1) 3.3mm OD x 18cm 12° BENT TUNNELER</li> <li>(1) Tunneler Sleeve</li> <li>(1) Dilator: (14F Sets) 4.7mm OD x 0.99mm ID x 15cm (14F) DILATOR, (16F Sets) (1) 5.4mm OD x 0.99mm ID x 15cm (16F) DILATOR</li> <li>(1) 6.2mm OD x 0.99mm ID x 15cm (18F) DILATOR</li> <li>(1) Valved Peelable Introducer: (14F Sets) 5.4mm ID x 19cm (16F) VALVED PEELABLE INTRODUCER, (16F Sets) (1) 5.7mm ID x 19cm (17F) VALVED PEELABLE INTRODUCER</li> <li>(1) Arterial Extension Set</li> <li>(1) Venous Extension Set</li> <li>(1) Removable Hub</li> <li>(3) Clamps</li> <li>(2) End Caps</li> <li>(1) Patient ID Card</li> <li>(1) Patient Information Packet</li> </ul>
32cm, 36cm, and 40cm length Set	<ul style="list-style-type: none"> <li>(1) Catheter</li> <li>(1) 1.3mm OD x 1.0mm ID x 70mm (18GA) INTRODUCER NEEDLE</li> <li>(2) 0.97mm x 100cm (.038) GUIDEWIRE J (R 3mm) TIP</li> <li>(2) Advancers</li> <li>(1) 3.3mm OD x 15cm 12° BENT Raulerson Ring Handled Y Adaptor Tunneler</li> <li>(1) 3.3mm OD x 18cm 12° BENT TUNNELER</li> <li>(1) Dilator: (14F Sets) 4.7mm OD x 0.99mm ID x 15cm (14F) DILATOR, (16F Sets) 5.4mm OD x 0.99mm ID x 15cm (16F) DILATOR</li> <li>(1) 6.2mm OD x 0.99mm ID x 15cm (18F) DILATOR</li> <li>(1) Introducer: (14F Sets) 5.4mm ID x 19cm (16F) VALVED PEELABLE INTRODUCER, (16F Sets) 5.7mm ID x 19cm (17F) VALVED PEELABLE INTRODUCER</li> <li>(1) Arterial Extension Set</li> <li>(1) Venous Extension Set</li> <li>(1) Removable Hub</li> <li>(3) Clamps</li> <li>(2) End Caps</li> <li>(1) Patient ID Card</li> <li>(1) Patient Information Packet</li> </ul>

Configuration Type	Kit Components
Set w/ Stylets	<p>(1) Catheter  (1) Stylet: 1.4mm OD 1.05mm ID x 417mm (24CM) STYLET, 1.4mm OD 1.05mm ID x 457mm (28CM) STYLET, 1.4mm OD 1.05mm ID x 497mm (32CM) STYLET, 1.4mm OD 1.05mm ID x 537mm (36CM) STYLET, 1.4mm OD 1.05mm ID x 577mm (40CM) STYLET  (1) 1.3mm OD x 1.0mm ID x 70mm (18GA) INTRODUCER NEEDLE (1) 0.98mm x 120cm (.038) GUIDEWIRE  (1) 0.97mm x 70cm (.038) GUIDEWIRE J (R 3mm) TIP (24, 28, 32, 36cm Sets)  (1) 0.91mm x 150cm (.035) GUIDEWIRE J (R 3mm) TIP (40cm Sets)  (2) Advancer</p> <p>(1) 3.3mm OD x 15cm 12° BENT Raulerson Ring Handled Y Adaptor Tunneler  (1) 3.3mm OD x 18cm 12° BENT TUNNELER  (1) Tunneler Sleeve  (1) 4.7mm OD x 0.99mm ID x 15cm (14F) DILATOR  (1) 6.2mm OD x 0.99mm ID x 15cm (18F) DILATOR  (1) 5.4mm ID x 19cm (16F) VALVED PEELABLE INTRODUCER  (1) Arterial Extension Set  (1) Venous Extension Set  (1) Removable Hub  (3) Clamps  (2) End Caps  (1) Patient ID Card  (1) Patient Information Packet</p>
Sheathless Technique Set	<p>(1) Catheter  (1) Stylet: 1.4mm OD 1.05mm ID x 417mm (24CM) STYLET, 1.4mm OD 1.05mm ID x 457mm (28CM) STYLET, 1.4mm OD 1.05mm ID x 497mm (32CM) STYLET, 1.4mm OD 1.05mm ID x 537mm (36CM) STYLET, 1.4mm OD 1.05mm ID x 577mm (40CM) STYLET  (1) 1.3mm OD x 1.0mm ID x 70mm (18GA) INTRODUCER NEEDLE  (1) 0.98mm x 120cm (.038) GUIDEWIRE  (1) Advancer  (1) 3.3mm OD x 15cm 12° BENT Raulerson Ring Handled Y Adaptor Tunneler  (1) 4.7mm OD x 0.99mm ID x 15cm (14F) DILATOR  (1) 6.2mm OD x 0.99mm ID x 15cm (18F) DILATOR  (1) Arterial Extension Set  (1) Venous Extension Set  (1) Removable Hub  (3) Clamps  (2) End Caps  (1) Patient ID Card  (1) Patient Information Packet</p>

## 2. Intended use of the device

Intended purpose	Split Stream® Catheters are intended for use in adult patients who do not have functional permanent vascular access or are not candidates for permanent vascular access for whom central venous vascular access for hemodialysis is deemed necessary based on the direction of a qualified, licensed physician. The catheter is intended to be used under the regular review and assessment of qualified health professionals. This catheter is for single use only.
Indication(s)	Split Stream® Catheters are indicated for short-term or long-term use where vascular access is required for 14 days or more for the purpose of hemodialysis.
Target population(s)	Split Stream® Catheters are intended for use in adult patients who do not have functional permanent vascular access or are not candidates for permanent vascular access for whom central venous vascular access for hemodialysis is deemed necessary based on the direction of a qualified, licensed physician. The catheter is not intended for use in pediatric patients.
Contraindications and/or limitations	<ul style="list-style-type: none"> <li>• Known or suspected allergies to any of the components of the catheter or the kit.</li> <li>• This device is contraindicated for patients exhibiting severe, uncontrolled coagulopathy or thrombocytopenia.</li> </ul>

## 3. Device description



Figure 1: Split Stream Catheter

Description of device	The Split Stream® Catheter is a long-term double lumen, single access catheter that is used to remove and return blood through two separate passages (lumens). The lumens are connected to two extensions by means of detachable adapter style hubs. Priming volumes are printed on each lumen. A polyester cuff is placed on the catheter's lumen for tissue ingrowth to anchor the catheter. The catheter incorporates Barium Sulphate to facilitate visualization under fluoroscopy or Xray. The catheter has been tested at flow rates of up to 500 mL/min. The catheter is available in a variety of sizes to accommodate physician preference and clinical needs.
-----------------------	---

Materials / substances in contact with patient tissue

The percentage ranges in the table below are based on the weights of the 24cm catheter (7.09g) and the 40cm catheter (11.58g).

14F Split Stream®	
Material	% Weight (w/w)
Polyurethane	56.22 - 59.41
Acetal co-polymer	15.01 - 17.33
Silicone	5.31 - 6.13
Barium sulfate	6.31 - 8.14
Stainless Steel	2.65 - 3.06
Nylon	4.30 - 4.96
Acrylonitrile Butadiene Styrene	3.52 - 4.07
Polyethylene terephthalate	1.65 - 1.91

The percentage ranges in the table below are based on the weights of the 24cm catheter (17.25g) and the 40cm catheter (19.92g).

16F Split Stream®	
Material	% Weight (w/w)
Polyurethane	56.46 - 59.51
Acetal co-polymer	14.93 - 17.15
Silicone	5.28 - 6.07
Barium sulfate	6.45 - 8.20
Stainless Steel	5.28 - 6.07
Nylon	4.27 - 4.91
Acrylonitrile Butadiene Styrene	3.51 - 4.03
Polyethylene terephthalate	1.65 - 1.90

**Note:** Per the instructions for use, the device is contraindicated for patients with known or suspected allergies to the above materials.

**Note:** Accessories containing stainless steel may contain up to 4% weight of the CMR substance cobalt.

Information on medicinal substances in the device

N/A

How the device achieves its intended mode of action

Hemodialysis catheters are centrally placed access tubes. A typical hemodialysis catheter uses a thin, flexible tube. The tube has two openings. The tube goes into a large vein. The vein is usually the internal jugular vein. Blood withdraws through one lumen of the catheter. The blood flows to the dialysis machine through a separate tubing set. The blood is then processed and filtered. The blood returns to the patient through the second lumen. This device is used when dialysis must start at once. Patients may not have a functioning AV fistula or graft. Catheter hemodialysis normally happens on a

	short-term basis. Long-term access may occur in some cases. For example, when there are problems supporting an AV fistula or graft.	
Sterilization Information	Contents sterile and non-pyrogenic in unopened, undamaged package. Sterilized by Ethylene Oxide.	
Previous generations / variants	Name of previous generation	Differences from current device
	N/A	N/A
Accessories intended for use in combination with the Split Stream Catheter	Name of Accessory	Description of Accessory
	Guidewire	For general intravascular use to facilitate the selective placement of medical devices in the vessel anatomy.
	Guidewire Advancer	Aid for introduction of guidewire into target vein.
	Stylet	Assist in catheter insertion
	Introducer Needle	Used for the percutaneous introduction of guidewires.
	Scalpel	A cutting device during surgical, pathology and minor medical procedures
	Tunneler	Instrument used to create a subcutaneous tunnel
	Tunneler Sleeve	Sleeve slides down the tunneler and over the catheter tip to secure the catheter to the tunneler.
	Peelable Introducer	Introducers are intended to obtain central venous access to facilitate catheter insertion into the central venous system.
	Dilator	Designed for percutaneous entry into a vessel in order to enlarge the opening of the vessel for the placement of a catheter in a vein.
	Removable Hub	The removable hub is intended to provide additional catheter securement and to minimize movement at the exit site.
End Cap	To keep clean and protect catheter luer between treatments.	
Other devices or products intended for use in combination with Split Stream Catheter	Name of Device or Product	Description of Device or Product
	Tegaderm	Adhesive wound dressing intended to protect the catheter from contamination when not in use
	Syringe	Attached to introducer needle to help capture blood return once introducer needle perforates targeted vein, prevent air embolism

#### 4. Risks and warnings

Residual risks and undesirable effects	<p>As per product IFU (IFU 40773BSI), All surgical procedures carry risk. Medcomp has implemented risk management processes to proactively find and mitigate these risks as far as possible without adversely affecting the benefit-risk profile of the device. After mitigation, residual risks and the possibility of adverse events from use of this product remain. Medcomp has determined that all residual risks are acceptable.</p>	
	<b>Residual Harm Type</b>	<b>Possible Adverse Events Associated with Harm</b>
	Bleeding	Bleeding (May be severe) Femoral Artery Bleed Hematoma Retroperitoneal Bleed
	Cardiac Event	Cardiac Arrhythmia Cardiac Tamponade
	Embolism	Air Embolus
	Infection	Bacteremia Endocarditis Exit Site Infection Septicemia Tunnel Infection
	Perforation	Inferior Vena Cava Puncture Laceration of the Vessel Perforation of the Vessel Pneumothorax Right Atrial Puncture Subclavian Artery Puncture Superior Vena Cava Puncture
	Thrombosis	Central Venous Thrombosis Lumen Thrombosis Subclavian Vein Thrombosis Vascular Thrombosis
	Miscellaneous Complications	Brachial Plexus Injury Femoral Nerve Damage Hemothorax Pleural Injury Thoracic Duct Laceration Venous Stenosis
	<b>Patient Residual Harm Category</b>	<b>Quantification of Residual Risks</b>
	<b>PMS Complaints (01 January 2016 – 31 March 2025)</b>	<b>PMCF Events</b>
	<b>Units Sold: 112,258</b>	<b>Units Studied: 107</b>
	<b>% of Devices</b>	<b>% of Devices</b>
Allergic Reaction	Not Reported	Not Reported
Bleeding	0.006%	Not Reported

	Cardiac Event	Not Reported	Not Reported
	Embolism	0.001%	Not Reported
	Infection	Not Reported	Not Reported
	Perforation	0.001%	Not Reported
	Stenosis	Not Reported	Not Reported
	Tissue Injury	Not Reported	Not Reported
	Thrombosis	Not Reported	Not Reported

Warnings and precautions	<p>All warnings have been reviewed against the risk analysis, PMS, and usability testing to validate consistency between the sources of information. As per product IFU (IFU 40773BSI), the Split Stream® Catheters have the following warnings:</p> <ul style="list-style-type: none"> <li>• Do not insert catheter in thrombosed vessels.</li> <li>• Do not advance the guidewire or catheter if unusual resistance is encountered.</li> <li>• Do not insert or withdraw the guidewire forcibly from any component. If the guidewire becomes damaged, guidewire and any associated componentry must be removed together.</li> <li>• Do not resterilize the catheter or accessories by any method.</li> <li>• Contents sterile and non-pyrogenic in unopened, undamaged package. STERILIZED BY ETHYLENE OXIDE</li> <li>• Do not re-use catheter or accessories as there may be a failure to adequately clean and decontaminate the device which may lead to contamination, catheter degradation, device fatigue, or endotoxin reaction.</li> <li>• Do not use catheter or accessories if package is opened or damaged.</li> <li>• Do not use catheter or accessories if any sign of product damage is visible or the use-by date has passed.</li> <li>• Do not use sharp instruments near the extension tubing or catheter lumen.</li> <li>• Do not use scissors to remove dressing.</li> <li>• Do not clamp over guidewire or stylet.</li> </ul> <p>Precautions listed in the Split Stream® Catheter IFU are as follows:</p> <ul style="list-style-type: none"> <li>• Examine catheter lumen and extensions before and after each treatment for damage.</li> <li>• To prevent accidents, ensure the security of all caps and bloodline connections prior to and between treatments.</li> <li>• Use only Luer Lock (threaded) Connectors with this catheter.</li> <li>• In the rare event that a hub or connector separates from any component during insertion or use, take all necessary steps and precautions to prevent blood loss or air embolism and remove the catheter.</li> <li>• Before attempting catheter insertion, ensure that you are familiar with the potential complications and their emergency treatment should any of them occur.</li> <li>• Repeated overtightening of bloodlines, syringes, and caps will reduce connector life and could lead to potential connector failure.</li> </ul>
--------------------------	---

- The catheter will be damaged if clamps other than what is provided with this kit are used.
- Avoid clamping near the Luer Lock and hub of the catheter. Clamping of the tubing repeatedly in the same location may weaken tubing.

Additional warnings and cautions listed in the Split Stream® Catheter IFU are as follows:

- Physician discretion is strongly advised when inserting this catheter in patients who are unable to take or hold a deep breath.
- Patients requiring ventilator support are at increased risk of pneumothorax during subclavian vein cannulation, which may cause complications.
- Extended use of the subclavian vein may be associated with subclavian vein stenosis.
- The incidence of infection may be increased with femoral vein insertion.
- Do not pull tunneler out at an angle. Keep tunneler straight to prevent damage to catheter tip.
- DO NOT grasp and pull the guidewire prior to releasing the J-Straightener. Damage to the guidewire may occur if it is pulled against the restraint of the J-Straightener.
- The length of the wire inserted is determined by the size of the patient. Monitor patient for arrhythmia throughout this procedure. The patient should be placed on a cardiac monitor during this procedure. Cardiac arrhythmias may result if guidewire is allowed to pass into the right atrium. The guidewire should be held securely during this procedure.
- Insufficient tissue dilation can cause compression of the catheter lumen against the guidewire causing difficulty in the insertion and removal of the guidewire from the catheter. This can lead to bending of the guidewire.
- The Valved Peelable Introducer is not designed for use in the arterial system or as a hemostatic device.
- DO NOT bend the sheath/dilator during insertion as bending will cause the sheath to prematurely tear. Hold the introducer close to the tip (approximately 3cm from tip) when initially inserting through the skin surface. To progress the introducer towards the vein, regrasp the introducer a few centimeters above the original grasp location and push down on the introducer. Repeat procedure until introducer is inserted to appropriate depth based on patient anatomy and physician's discretion.
- Never leave sheath in place as an indwelling catheter. Damage to the vein will occur.
- Assure that all air has been aspirated from the catheter and extensions. Failure to do so may result in air embolism.
- Failure to verify catheter placement may result in serious trauma or fatal complications.
- Care must be taken when using sharp objects or needles in close proximity to catheter lumen. Contact from sharp objects may cause catheter failure.
- Only clamp catheter with in-line clamps provided.

	<ul style="list-style-type: none"> <li>• Extension clamps should only be open for aspiration, flushing, and dialysis treatment.</li> <li>• Patients must not swim, shower, or soak dressing while bathing.</li> <li>• Always review hospital or unit protocol, potential complications and their treatment, warnings, and precautions prior to undertaking any type of mechanical or chemical intervention in response to catheter performance problems.</li> <li>• Only a physician familiar with the appropriate techniques should attempt the following procedures.</li> <li>• Due to the risk of exposure to HIV (Human Immunodeficiency Virus) or other blood borne pathogens, health care professionals should always use Universal Blood and Body Fluid Precautions in the care of all patients.</li> <li>• Do not pull distal end of catheter through incision as contamination of wound may occur.</li> </ul>
Other relevant aspects of safety (ex. field safety corrective actions, etc.)	For a period of 01 January 2020 to 31 March 2025 there were 285 complaints for 61,824 units sold, giving an overall complaint rate of 0.46%. No events resulted in recalls during the review period.

**5. Summary of clinical evaluation and post-market clinical follow-up (PMCF)**

Summary of clinical data related to the subject device			
The below table displays the device insertion case numbers identified and used for clinical performance evaluation in each clinical data source.			
Clinical Literature	PMCF Data	Total Cases	User Survey Responses
179	107	286	0
<p>Clinical performance was measured using parameters including but not limited to dwell time, catheter insertion outcomes, and adverse event rates. Critical clinical parameters extracted from these studies met standards set forth in the guidelines for the State of the Art. There were no unforeseen adverse events or other high occurrences of adverse events detected in any of the clinical activities.</p> <p>Medcomp® catheters are subjected to, and must pass, simulated use testing intended to replicate use 3 times per week for 12 months as part of device development. The Split Stream® Catheter passed this testing. Although Medcomp® catheters contain no materials which degrade over time, fully functional catheters may be removed for other reasons, such as intractable infection, change of therapy (such as Renal replacement (transplant) or use of an arterio-venous graft/fistula). Published clinical literature does not always focus on the physical lifetime of a catheter for these reasons. In the case of the Split Stream® Catheter, 45, 49, and 47 catheters had 112, 130.7, and 130.5 day mean durations of use that has been found in clinical use reported to date. Based on this information, the Split Stream® Catheter has a 12-month lifetime; however, the decision to remove and/or replace the catheter should be based on clinical performance and need, and not any predetermined point in time.</p>			

**Summary of clinical data related to the equivalent device (if applicable)**

Clinical evidence from published literature and PMCF activities has been generated specific to known and unknown variants of the subject device. The equivalency rationale in the updated clinical evaluation report will demonstrate that the clinical evidence available for these variants is representative of the range of device variants in the device family.

There are no clinical or biological differences between variants within the subject device family, and the potential impact of the technical differences will be rationalized in the updated clinical evaluation report.

**Summary of clinical data from pre-market investigations (if applicable)**

No pre-market clinical devices were used for the device's clinical evaluation.

**Summary of clinical data from other sources:**

**Source: Summary of Published Literature**

Clinical evidence literature searches have found three published literature articles representing 179 Split Stream® device family specific cases. The articles include one randomized controlled trial (Silva et al., 2008), one retrospective study (Leou et al., 2013), and one case study (Ogawa et al., 2021).

**Bibliography:**

Leou S, Garnier F, Testevuide P, et al. Évaluation des complications infectieuses liées aux cathéters veineux centraux d'hémodialyse en Polynésie française. *Néphrologie & Thérapeutique*. 2013;9(3):137-42.

Ogawa, Tomonari; Inamura, Megumi; Kawai, Yuichiro; Yamamoto, Ryo; Yasuda, Kunihiko; Shimizu, Taisuke; Tamaru, Jun-ichi; Hasegawa, Hajime; (2021). Difficulty removing dialysis cuff catheter after its adhesion to the right atrium #journal#, (#issue#), 1129729821993981

Silva J, Teixeira e C, Baptista A, Ramos A, Ponce P. Catheter-related bacteremia in hemodialysis: which preventive measures to take? *Nephron Clinical practice*. 2008;110(4):c251-257.

**Source: LTHD Data Collection Survey Report**

The Long-Term Hemodialysis Catheter Data Collection Survey was intended to gather safety and performance outcome information from sites that purchase Medcomp long-term hemodialysis catheters for use in EU MDR clinical evaluation. Responses were requested to be completed by physicians or other site employees with oversight and direction from a physician. The surveys were distributed globally to existing Medcomp customers. Responses were collected from twenty-one sites, spanning nine countries (Colombia, Croatia, El Salvador, Greece, Italy, Netherlands, Panama, Uruguay, and USA) across North America, South/Latin America, and Europe.

At least partial data was collected on 56 Split Stream® catheter product family cases. Of these 56 cases, all were described as 14F, with 29 catheters of 24cm length, 18 catheters of 28cm length, and 9 catheters of 32cm length. Information was collected on Insertion Success

(100%, n=56), but no dwell time information was provided. Due to the absence of dwell time information, rate information could not be derived, and only insertion success was concluded to be within State of the Art safety and performance outcome measures from published literature.

**Source: PMCF\_Infusion\_211**

The Infusion Product Line Data Collection Survey aimed to assess safety and performance outcome information for all variants of Medcomp Infusion Ports, PICCs, Midlines, and CVCs. 70 survey responses were collected from 17 countries representing 471 device cases.

7 Split Stream® cases, all described as 14F, inclusive of several variant devices across length (28cm, 32cm, 40cm) were collected. The following outcome measures were collected for Medcomp Split Stream® devices:

- Procedural Outcomes – 100%
- Catheter Related Blood Stream Infection – No Events Reported
- Catheter Associated Venous Thrombus – No Events Reported
- Exit Site Infection – No Events Reported

**Source: PMCF\_Medcomp\_211**

The Medcomp User Survey acquired responses from healthcare personnel familiar with any number of Medcomp's product offerings.

28 respondents responded that they or their facility have used Medcomp long-term hemodialysis catheters; there were no respondents that have used the Split Stream® device. There were no differences in mean user sentiments within long-term hemodialysis catheters across State of the Art Performance and Safety Outcome Measures or between device types relating to safety or performance.

The following data points were collected from users of Medcomp long-term hemodialysis catheters (n=28):

- (Mean Likert Scale Response) Catheters function as intended – 4.8 / 5
- (Mean Likert Scale Response) Packaging allows for aseptic presentation – 4.8 / 5
- (Mean Likert Scale Response) Benefit outweighs the risk – 4.7 / 5
- Dwell Time (n=26) – 167 days (**95%CI: 130 – 203**)

**Source: PMCF\_LTHD\_242**

The Long-Term Hemodialysis (LTHD) Truveta data analysis assessed safety and performance outcome information for Medcomp® and competitor devices present in Truveta Studio. Truveta data comes from a growing collective of more than 30 health systems that provide 17% of the daily clinical care across all 50 U.S. states from 800 hospitals and 20,000 clinics, representing the full diversity of the United States. The population used for data analysis was derived utilizing Truveta Studio's proprietary coding language (Prose) and unique device identifier (UDI) codes representing all saleable Medcomp® LTHD devices and LTHD devices distributed and/or manufactured by other companies.

44 Split Stream® cases inclusive of several variant devices were collected. All cases were described as 14F and Straight Cases, configurations (straight), and lengths (24cm, 32cm),

representation of 24cm and 32cm length catheters. The following State of the Art safety and performance outcome measures were observed for Medcomp Split Stream® devices:

- Catheter Related Blood Stream Infection – 0 per 1,000 catheter days (95%CI: 0 – 0.93)
- Catheter Associated Venous Thrombus – 0 per 1,000 catheter days (95%CI: 0 – 0.93)
- Exit Site Infection – 0 per 1,000 catheter days (95%CI: 0 – 0.93)
- Tunnel Infection – 0 per 1,000 catheter days (95%CI: 0 – 0.93)
- Dwell Time – 82.5 days (95%CI: 0 – 199.04)

The catheter brand logistic regression model did not find that any Medcomp® catheter brands were statistically significantly associated with an increase of the incidence of CRBSI. The brand agnostic logistic regression found that pediatric age group (0–19 years), femoral vein insertion site, catheters that were the fourth or beyond in sequence for a given patient, split-tip designs, and pre-curved configurations were statistically significantly associated with the incidence of CRBSI. The Split Cath® III was associated with a statistically significant decrease in CRBSI incidence in the brand model (OR: 0.46 95%CI: 0.33 - 0.63), and both shorter catheter length (<=24cm) and smaller French size (<14.5F) in the brand agnostic model.

**Overall summary of clinical safety and performance**

Upon review of the Split Stream® Catheter data across all sources, it is possible to conclude that the benefits of the subject device, which is facilitating hemodialysis in patients in whom other therapies or conservative care are not indicated or desirable as determined by the physician, outweigh the overall and individual risks when the device is used as intended by the manufacturer. It is the manufacturer’s and clinical expert evaluator’s opinion that activities both complete and ongoing are sufficient to support the safety, efficacy, and acceptable benefit/risk profile of the Split Stream® catheters.

Outcome	Benefit/Risk Acceptability Criteria	Desired Trend	Clinical Literature (Subject Device)	PMCF Data (Subject Device)
<b>Performance</b>				
Dwell Time	Greater than 40 days	↑	105.4 days - 130.7 days (Summary of Published Literature)	82.5 days (PMCF_LTHD_242)
Procedural Outcomes	Greater than 93.3%	↑	ND*	100% (LTHD Data Collection Survey Report & PMCF_Infusion_211)
<b>Safety</b>				
Catheter Related Blood Stream Infection (CRBSI)	Less than 4.8 incidents of CRBSI per 1,000 catheter days	↓	0.16 – 1.78 per 1,000 catheter days (Summary of Published Literature)	No Events Reported (LTHD Data Collection Survey Report & PMCF_Infusion_211)

				0 per 1,000 catheter days <b>(PMCF_LTHD_242)</b>
Tunnel Infection Rate	Less than 2.8 incidents of tunnel infection per 1,000 catheter days	↓	0.25** - 0.59** per 1,000 catheter days <b>(Summary of Published Literature)</b>	No Events Reported <b>(LTHD Data Collection Survey Report &amp; PMCF_Infusion_211)</b>  0 per 1,000 catheter days <b>(PMCF_LTHD_242)</b>
Exit Site Infection Rate	Less than 3.2 incidents of exit site infection per 1,000 catheter days	↓	0.25** - 0.59** per 1,000 catheter days <b>(Summary of Published Literature)</b>	No Events Reported <b>(LTHD Data Collection Survey Report &amp; PMCF_Infusion_211)</b>  0 per 1,000 catheter days <b>(PMCF_LTHD_242)</b>
Catheter Associated Venous Thrombus (CAVT)	Less than 3.04 incidents of CAVT per 1,000 catheter days	↓	2.13*** - 3.2*** per 1,000 catheter days <b>(Summary of Published Literature)</b>	No Events Reported <b>(LTHD Data Collection Survey Report &amp; PMCF_Infusion_211)</b>  0 per 1,000 catheter days <b>(PMCF_LTHD_242)</b>

\*ND = No data on the clinical parameter

\*\*Silva et al., 2008 and Leou et al., 2013 do not differentiate between exit site and tunnel infection

\*\*\*Obstruction episodes from Silva et al., 2008 is used to approximate the rate of CAVT

**On-going or planned Post-Market Clinical Follow-up (PMCF)**

<b>Activity</b>	<b>Description</b>	<b>Reference</b>	<b>Timeline</b>
Multi-center Patient-Level Case Series	Collect additional clinical data on the device by acquiring case data healthcare personnel familiar with the device.	PMCF_LTHD_241	Q4 2025
State of the Art Literature Search	Identify risks and trends with use of	SAP-HD	Q2 2026

	similar devices by reviewing applicable standards, published literature, conference abstracts, guidance documents and recommendations; information relating to the medical condition managed by the device and medical alternatives available for the same target treated population.		
Clinical Evidence Literature Search	Identify risks and trends with use of the device by reviewing any clinical data relevant to the device from published literature.	LRP-HD	Q2 2026
Global Trial Database Search	Identify ongoing clinical trials involving Split Stream® catheters.	N/A	Q2 2026
No emerging risks, complications or unexpected device failures have been detected from PMCF activities.			

## 6. Possible therapeutic alternatives

The Kidney Disease Outcomes Quality Initiative (KDOQI) 2019 clinical practice guidelines have been used to support the below recommendations for treatments.

Therapy	Benefits	Disadvantages	Key Risks
AV Fistula	<ul style="list-style-type: none"> <li>Permanent vascular access solution</li> <li>Lower complication rate than hemodialysis via catheter</li> </ul>	<ul style="list-style-type: none"> <li>Requires time to mature</li> <li>Patients must sometimes self-cannulate</li> </ul>	<ul style="list-style-type: none"> <li>Stenosis</li> <li>Thrombosis</li> <li>Aneurysm</li> <li>Pulmonary hypertension</li> <li>Steal Syndrome</li> <li>Septicemia</li> </ul>
Hemodialysis Catheter	<ul style="list-style-type: none"> <li>Useful for quick vascular access without AV Fistula in place</li> <li>Can be used as a bridge dialysis</li> </ul>	<ul style="list-style-type: none"> <li>Not a permanent solution</li> <li>Catheter dysfunction can disrupt regular treatment</li> </ul>	<ul style="list-style-type: none"> <li>Post-procedural bleeding</li> <li>Infection</li> <li>Thrombosis</li> <li>Decreased blood flow in</li> </ul>

Therapy	Benefits	Disadvantages	Key Risks
	method between other therapies	<ul style="list-style-type: none"> <li>Benefit is not equal for all patient populations</li> </ul>	dysfunctional catheter <ul style="list-style-type: none"> <li>Cardiovascular events</li> <li>Fibrin sheath formation around catheter</li> <li>Septicemia</li> </ul>
Peritoneal Dialysis	<ul style="list-style-type: none"> <li>Less restrictive diet than hemodialysis</li> <li>Does not require hospitalization, can be done in any clean place</li> </ul>	<ul style="list-style-type: none"> <li>Clearance of impurities is limited by dialysate flow and peritoneal area</li> </ul>	<ul style="list-style-type: none"> <li>Peritonitis</li> <li>Septicemia</li> <li>Fluid overload</li> </ul>
Kidney Transplant	<ul style="list-style-type: none"> <li>Better quality of life compared to HD</li> <li>Lower risk of death compared to HD</li> <li>Fewer dietary restrictions compared to HD</li> </ul>	<ul style="list-style-type: none"> <li>Requires a donor which can take time</li> <li>More risky for certain groups (aged, diabetics, etc.)</li> <li>Patient must take rejection medication for life               <ul style="list-style-type: none"> <li>Rejection medication has side effects</li> </ul> </li> </ul>	<ul style="list-style-type: none"> <li>Thrombosis</li> <li>Hemorrhage</li> <li>Ureteral blockage               <ul style="list-style-type: none"> <li>Infection</li> </ul> </li> <li>Organ rejection               <ul style="list-style-type: none"> <li>Death</li> <li>Myocardial infarction</li> <li>Stroke</li> </ul> </li> </ul>
Comprehensive Conservative Care	<ul style="list-style-type: none"> <li>Less imposed symptom burden than dialysis</li> <li>Preserves life satisfaction</li> </ul>	<ul style="list-style-type: none"> <li>May aggravate clinical condition</li> <li>Not designed to treat, but to minimize adverse events</li> </ul>	<ul style="list-style-type: none"> <li>Treatment may not actually minimize risks associated with CKD</li> </ul>

## 7. Suggested profile and training for users

The catheter should be inserted, manipulated, and removed by a qualified, licensed physician or other qualified health care professional under the direction of a physician. In certain circumstances, patients who may be suitable for home hemodialysis may manipulate the external connections of the catheter.

As per guidelines stated from the International Society of Hemodialysis, if home dialysis is recommended, each patient will undergo a thorough training in order to obtain optimal results from home dialysis treatments. The objectives of the training program are to (1) provide the appropriate amount of information to ensure that the patient will be able to dialyze safely at home; (2) enable the patient to monitor and manage other elements of his or her chronic kidney disease, such as obtaining samples for lab work and maintaining appropriate nutrition and diet; and (3) help the patient and his or her care partner(s) cope with barriers and fears associated

with home HD During training, the patient will also receive technical education on the operations and maintenance of the water treatment system.

During training, the ideal nurse trainer-to-patient ratio is typically 1:1. An idealized schedule of training is created, with weekly areas of focus and training objectives. In practice, however, training is individualized to address any identified learning barriers or risks for failure.

### 8. Reference to any harmonized standards and Common Specifications (CS) applied

Harmonized Standard or CS	Revision	Title or Description	Level of Compliance
EN ISO 14971	2019	Medical devices. Application of risk management to medical devices	Full
EN ISO 10555-1	2023	Intravascular catheters. Sterile and single-use catheters. General requirements	Full
ISO 10555-3	2013	Intravascular catheters. Sterile and single-use catheters. Central venous catheters	Full
EN ISO 11607-1	2020+A1:2023	Packaging for terminally sterilized medical devices. Requirements for materials, sterile barrier systems and packaging systems	Full
EN ISO 11607-2	2020+A1:2023	Packaging for terminally sterilized medical devices. Validation requirements for forming, sealing and assembly processes	Full
MEDDEV 2.7/1	Rev 4	Clinical Evaluation: A Guide for Manufacturers and Notified Bodies Under Directives 93/42/EEC and 90/385/EEC	Full
MEDDEV 2.12/2	Rev. 2	GUIDELINES ON MEDICAL DEVICES POST MARKET CLINICAL FOLLOW-UP STUDIES A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES	Full
EN ISO 14155	2020	Clinical investigation of medical devices for human subjects — Good clinical practice	Full
MDCG 2020-6	2020	Clinical evidence needed for medical devices previously CE marked under Directives 93/42/EEC or 90/385/EEC	Full
MDCG 2020-7	2020	Post-market clinical follow-up (PMCF) Plan Template A guide for manufacturers and notified bodies	Full
MDCG 2020-8	2020	Post-market clinical follow-up (PMCF) Evaluation Report Template A guide for manufacturers and notified bodies	Full
MDCG 2022-9	2022	Summary of safety and clinical performance	Full
MDCG 2022-21	2022	Guidance on Periodic Safety Update Report (PSUR) According to Regulation EU 2017/745 (MDR)	Full

Harmonized Standard or CS	Revision	Title or Description	Level of Compliance
ISO 10993-1	2020	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process	Full
ISO 10993-18	2020+A1:2023	Biological evaluation of medical devices — Part 18: Chemical characterization of medical device materials within a risk management process	Full
EN ISO 10993-7	2008 + A1: 2022	Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals — Amendment 1: Applicability of allowable limits for neonates and infants	Full
EN ISO 11135	2014 + A1: 2019	Sterilization of health-care products. Ethylene oxide. Requirements for the development, validation and routine control of a sterilization process for medical devices	Full
ISO 14644-1	2015	Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration	Full
ISO 14644-2	2015	Cleanrooms and associated controlled environments — Part 2: Monitoring to provide evidence of cleanroom performance related to air cleanliness by particle concentration	Full
EN 556-1	2024	Sterilization of medical devices. Requirements for medical devices to be designated "STERILE". Requirements for terminally sterilized medical devices	Full
EN ISO 11737-1	2018 + A1: 2021	Sterilization of health care products. Microbiological methods. Determination of a population of microorganisms on products	Full
EN 11737-3	2023	Sterilization of health care products. Microbiological methods — Bacterial endotoxin testing	Full
ANSI/AAMI ST72	2019	Bacterial endotoxins-Test methods, routine monitoring, and alternatives to batch testing	Full
EN ISO 20417	2021	Medical Devices - Information supplied by the manufacturer	Full
EN ISO 15223-1	2021	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements	Full

Harmonized Standard or CS	Revision	Title or Description	Level of Compliance
EN 62366-1	2015 + A1: 2020	Medical devices — Part 1: Application of usability engineering to medical devices	Full
ASTM D4332	2022	Standard Practice for Conditioning Containers, Packages, or Packaging Components for Testing	Full
ASTM F2503	2023e1	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment	Full
EN ISO 11070	2014+A1:2018	Sterile single-use intravascular introducers, dilators and guidewires	Full
ISO 594-1	1986	Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements	Full
ISO 594-2	1998	Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock Fittings	Full
ASTM D4169	2023e1	Standard Practice for Performance Testing of Shipping Containers and Systems	Full
EN ISO 13485	2016 + A11: 2021	Medical Devices – Quality Management system – Requirements for Regulatory Purposes	Full
PD CEN ISO/TR 20416	2020	Medical devices — Post-market surveillance for manufacturers	Full
EN ISO 80369-7	2021	Small-bore connectors for liquids and gases in healthcare applications Part 7: Connectors for intravascular or hypodermic applications	Full
MDCG 2018-1	Rev. 4	Guidance on BASIC UDI-DI and changes to UDI-DI	Full
EN ISO 11140-1	2014	Sterilization of health care products — Chemical indicators Part 1: General requirements	Full
EN ISO/IEC 17025	2017	General requirements for the competence of testing and calibration laboratories	Full
Regulation (EU) 2017/745	2017	Regulation (EU) 2017/745 of the European Parliament and of the Council	Full

---

## PATIENTS

---

### SUMMARY OF SAFETY AND CLINICAL PERFORMANCE

Revision: SSCP-007 Rev. 5

Date: 05SEP2025

This Summary of Safety and Clinical Performance (SSCP) is intended to provide public access to an updated summary of the main aspects of the safety and clinical performance of the device. The information presented below is intended for patients or lay persons. A more extensive summary of safety and clinical performance prepared for healthcare professionals is found in the first part of this document.

---

### IMPORTANT INFORMATION

The SSCP is not intended to give general advice on the treatment of a medical condition. Please contact your healthcare professional in case you have questions about your medical condition or about the use of the device in your situation.

This SSCP is not intended to replace an Implant Card or the Instructions for Use to provide information on the safe use of the device.

---

#### 1. Device identification and general information

Device trade name(s)	Split Stream®
Manufacturer name and address	Medical Components, Inc. 1499 Delp Drive Harleysville, PA 19438 USA
Basic UDI-DI	00884908249NH
Date first CE certificate was issued for this device	2003

The devices in scope of this document are all long-term hemodialysis catheter sets. The device part numbers are organized into variant categories. These devices are distributed as procedure trays. Procedure trays come in different configurations.

Variant Devices:

Variant Description	Part Number
14F x 24cm Split Stream	10298-824
	10028-824
14F x 28cm Split Stream	10298-828
	10028-828

Variant Description	Part Number
14F x 32cm Split Stream	10298-832 10028-832
14F x 36cm Split Stream	10298-836 10028-836
14F x 40cm Split Stream	10298-840 10028-840
16F x 24cm Split Stream	10062-824
16F x 28cm Split Stream	10062-828
16F x 32cm Split Stream	10062-832
16F x 36cm Split Stream	10062-836
16F x 40cm Split Stream	10062-840

Procedure Trays:

Catalog Code	Part Number	Description
S24SE.	10298-824	14F x 24cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S28SE.	10298-828	14F x 28cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S32SE.	10298-832	14F x 32cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S36SE.	10298-836	14F x 36cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
S40SE.	10298-840	14F x 40cm Split Stream® Catheter Sheathless Technique Set (Cuff 19cm From Tip)
SST24SE.	10298-824	14F x 24cm Split Stream® Catheter w/ Stylet Set (Cuff 19cm From Tip)
SST28SE.	10298-828	14F x 28cm Split Stream® Catheter w/ Stylet Set (Cuff 23cm From Tip)
SST32SE.	10298-832	14F x 32cm Split Stream® Catheter w/ Stylet Set (Cuff 27cm From Tip)
SST36SE.	10298-836	14F x 36cm Split Stream® Catheter w/ Stylet Set (Cuff 31cm From Tip)
SST40SE.	10298-840	14F x 40cm Split Stream® Catheter w/ Stylet Set (Cuff 35cm From Tip)
SST24E.	10028-824	14F x 24cm Split Stream® Catheter Set (Cuff 19cm From Tip)
SST28E.	10028-828	14F x 28cm Split Stream® Catheter Set (Cuff 23cm From Tip)
SST32E.	10028-832	14F x 32cm Split Stream® Catheter Set (Cuff 27cm From Tip)
SST36E.	10028-836	14F x 36cm Split Stream® Catheter Set (Cuff 31cm From Tip)
SST40E.	10028-840	14F x 40cm Split Stream® Catheter Set (Cuff 35cm From Tip)
SST2416E.	10062-824	16F x 24cm Split Stream® Catheter Set (Cuff 19cm From Tip)
SST2816E.	10062-828	16F x 28cm Split Stream® Catheter Set (Cuff 23cm From Tip)
SST3216E.	10062-832	16F x 32cm Split Stream® Catheter Set (Cuff 27cm From Tip)
SST3616E.	10062-836	16F x 36cm Split Stream® Catheter Set (Cuff 31cm From Tip)
SST4016E.	10062-840	16F x 40cm Split Stream® Catheter Set (Cuff 35cm From Tip)

Configurations of Procedure Trays:

Configuration Type
24cm and 28cm length Sets
32cm, 36cm, and 40cm length Set
Set w/ Stylets
Sheathless Technique Set

## 2. Intended use of the device

Intended purpose	Split Stream® Catheters are intended for use in adult patients who do not have functional permanent vascular access or are not candidates for permanent vascular access for whom central venous vascular access for hemodialysis is deemed necessary based on the direction of a qualified, licensed physician. The catheter is intended to be used under the regular review and assessment of qualified health professionals. This catheter is for single use only.
Indication(s)	Split Stream® Catheters are indicated for short-term or long-term use where vascular access is required for 14 days or more for the purpose of hemodialysis.
Intended patient group(s)	Split Stream® Catheters are intended for use in adult patients who do not have functional permanent vascular access or are not candidates for permanent vascular access for whom central venous vascular access for hemodialysis is deemed necessary based on the direction of a qualified, licensed physician. The catheter is not intended for use in pediatric patients.
Contraindications	<ul style="list-style-type: none"> <li>Known or suspected allergies to any of the components of the catheter or the kit.</li> <li>This device is contraindicated for patients exhibiting severe, uncontrolled coagulopathy or thrombocytopenia.</li> </ul>

## 3. Device description



**Figure 1: Split Stream Catheter**

Description of device	The Split Stream® Catheters are long-term catheters. The catheters are double tubed. The catheters remove and return blood through two separate lines. The tubes are assembled during catheter placement. The tubes list the priming volume. A polyester cuff on the catheter tubing helps attach the catheter to the patient.
-----------------------	--

Materials / substances in contact with patient tissue	<p>The percentage ranges below are based on catheter weights. The 24cm catheter weighs 7.09 grams. The 40cm catheter weighs 11.58 grams.</p>																				
	<table border="1" data-bbox="625 325 1328 739"> <thead> <tr> <th colspan="2" data-bbox="625 325 1328 367">14F Split Stream®</th> </tr> <tr> <th data-bbox="625 367 1068 409">Material</th> <th data-bbox="1068 367 1328 409">% Weight (w/w)</th> </tr> </thead> <tbody> <tr> <td data-bbox="625 409 1068 451">Polyurethane</td> <td data-bbox="1068 409 1328 451">56.22 - 59.41</td> </tr> <tr> <td data-bbox="625 451 1068 493">Acetal co-polymer</td> <td data-bbox="1068 451 1328 493">15.01 - 17.33</td> </tr> <tr> <td data-bbox="625 493 1068 535">Silicone</td> <td data-bbox="1068 493 1328 535">5.31 - 6.13</td> </tr> <tr> <td data-bbox="625 535 1068 577">Barium sulfate</td> <td data-bbox="1068 535 1328 577">6.31 - 8.14</td> </tr> <tr> <td data-bbox="625 577 1068 619">Stainless Steel</td> <td data-bbox="1068 577 1328 619">2.65 - 3.06</td> </tr> <tr> <td data-bbox="625 619 1068 661">Nylon</td> <td data-bbox="1068 619 1328 661">4.30 - 4.96</td> </tr> <tr> <td data-bbox="625 661 1068 703">Acrylonitrile Butadiene Styrene</td> <td data-bbox="1068 661 1328 703">3.52 - 4.07</td> </tr> <tr> <td data-bbox="625 703 1068 745">Polyethylene terephthalate</td> <td data-bbox="1068 703 1328 745">1.65 - 1.91</td> </tr> </tbody> </table>	14F Split Stream®		Material	% Weight (w/w)	Polyurethane	56.22 - 59.41	Acetal co-polymer	15.01 - 17.33	Silicone	5.31 - 6.13	Barium sulfate	6.31 - 8.14	Stainless Steel	2.65 - 3.06	Nylon	4.30 - 4.96	Acrylonitrile Butadiene Styrene	3.52 - 4.07	Polyethylene terephthalate	1.65 - 1.91
	14F Split Stream®																				
Material	% Weight (w/w)																				
Polyurethane	56.22 - 59.41																				
Acetal co-polymer	15.01 - 17.33																				
Silicone	5.31 - 6.13																				
Barium sulfate	6.31 - 8.14																				
Stainless Steel	2.65 - 3.06																				
Nylon	4.30 - 4.96																				
Acrylonitrile Butadiene Styrene	3.52 - 4.07																				
Polyethylene terephthalate	1.65 - 1.91																				
<p>The percentage ranges below are based on catheter weights. The 24cm catheter weighs 17.25 grams. The 40cm catheter weighs 19.92 grams.</p>																					
<table border="1" data-bbox="625 903 1328 1316"> <thead> <tr> <th colspan="2" data-bbox="625 903 1328 945">16F Split Stream®</th> </tr> <tr> <th data-bbox="625 945 1068 987">Material</th> <th data-bbox="1068 945 1328 987">% Weight (w/w)</th> </tr> </thead> <tbody> <tr> <td data-bbox="625 987 1068 1029">Polyurethane</td> <td data-bbox="1068 987 1328 1029">56.46 - 59.51</td> </tr> <tr> <td data-bbox="625 1029 1068 1071">Acetal co-polymer</td> <td data-bbox="1068 1029 1328 1071">14.93 - 17.15</td> </tr> <tr> <td data-bbox="625 1071 1068 1113">Silicone</td> <td data-bbox="1068 1071 1328 1113">5.28 - 6.07</td> </tr> <tr> <td data-bbox="625 1113 1068 1155">Barium sulfate</td> <td data-bbox="1068 1113 1328 1155">6.45 - 8.20</td> </tr> <tr> <td data-bbox="625 1155 1068 1197">Stainless Steel</td> <td data-bbox="1068 1155 1328 1197">5.28 - 6.07</td> </tr> <tr> <td data-bbox="625 1197 1068 1239">Nylon</td> <td data-bbox="1068 1197 1328 1239">4.27 - 4.91</td> </tr> <tr> <td data-bbox="625 1239 1068 1281">Acrylonitrile Butadiene Styrene</td> <td data-bbox="1068 1239 1328 1281">3.51 - 4.03</td> </tr> <tr> <td data-bbox="625 1281 1068 1323">Polyethylene terephthalate</td> <td data-bbox="1068 1281 1328 1323">1.65 - 1.90</td> </tr> </tbody> </table> <p><b>Note:</b> The device should not be used if you are allergic to the above materials.</p> <p><b>Note:</b> Accessories containing stainless steel may contain up to 4% weight of the CMR substance cobalt.</p>	16F Split Stream®		Material	% Weight (w/w)	Polyurethane	56.46 - 59.51	Acetal co-polymer	14.93 - 17.15	Silicone	5.28 - 6.07	Barium sulfate	6.45 - 8.20	Stainless Steel	5.28 - 6.07	Nylon	4.27 - 4.91	Acrylonitrile Butadiene Styrene	3.51 - 4.03	Polyethylene terephthalate	1.65 - 1.90	
16F Split Stream®																					
Material	% Weight (w/w)																				
Polyurethane	56.46 - 59.51																				
Acetal co-polymer	14.93 - 17.15																				
Silicone	5.28 - 6.07																				
Barium sulfate	6.45 - 8.20																				
Stainless Steel	5.28 - 6.07																				
Nylon	4.27 - 4.91																				
Acrylonitrile Butadiene Styrene	3.51 - 4.03																				
Polyethylene terephthalate	1.65 - 1.90																				
Information on medicinal substances in the device	N/A																				
How the device achieves its intended mode of action	<p>Hemodialysis catheters are centrally placed access tubes. A typical hemodialysis catheter uses a thin, flexible tube. The tube has two openings. The tube goes into a large vein. The vein is usually the internal jugular vein. Blood withdraws through one lumen of the catheter. The blood flows to the dialysis machine through a separate tubing set. The blood is then processed and filtered. The blood returns to the patient through the second lumen. This device is used when dialysis must start at once. Patients may not have a functioning AV</p>																				

	fistula or graft. Catheter hemodialysis normally happens on a short-term basis. Long-term access may occur in some cases. For example, when there are problems supporting an AV fistula or graft.	
Sterilization Information	Contents sterile and non-pyrogenic in unopened, undamaged package. Sterilized by Ethylene Oxide.	
Description of accessories	<b>Name of Accessory</b>	<b>Description of Accessory</b>
	<b>Guidewire</b>	Acts as a path for other components.
	<b>Guidewire Advancer</b>	Helps guidewire introduction.
	<b>Stylet</b>	Assist in catheter insertion.
	<b>Introducer Needle</b>	Placed into the target vein to gain access.
	<b>Tunneler</b>	Creates a pocket in between muscle and skin for catheter.
	<b>Tunneler Sleeve</b>	Helps secure the catheter to the tunneler.
	<b>Peelable Introducer</b>	Used to get central venous access.
	<b>End Cap</b>	To keep the catheter clean between treatments.
	<b>Removable Hub</b>	Provides additional catheter securement.
	<b>Dilator</b>	Used to make the opening of a vessel larger.
	<b>Scalpel</b>	A cutting device.
	<b>Syringe</b>	Helps get blood return once the needle punctures the vein.
<b>Tegaderm</b>	Dressing that protects the catheter from contamination.	

#### 4. Risks and warnings

Contact your healthcare professional if you believe you are experiencing side-effects related to the device or its use or if you are concerned about risks. This document does not replace a consultation with your healthcare professional if needed.

How potential risks have been controlled or managed	<p>There have been 61,824 devices sold since January 2020. There are side effects and risks associated with the device. These include:</p> <ul style="list-style-type: none"> <li>• Infection</li> <li>• Bleeding</li> <li>• Catheter Removal</li> <li>• Catheter Replacement</li> </ul> <p>These risks are reduced to an acceptable level. The labeling describes the risks. The benefit of the device is access for hemodialysis when alternatives are not suitable. These benefits outweigh the risks.</p>
---	---

<p>Remaining risks and undesirable effects</p>	<p>The Split Stream® catheter is associated with risks. These include:</p>																																					
	<ul style="list-style-type: none"> <li>• Procedural Delays</li> <li>• Thrombosis</li> <li>• Infections</li> <li>• Perforations</li> <li>• Embolism</li> <li>• Cardiac Event</li> <li>• Dissatisfaction</li> </ul>																																					
	<p>These risks are consistent with risks of other dialysis catheters. They are not unique to the Medcomp product. Some of the most common reactions include infection. Infection may be associated with general surgical procedure and hospitalization. Infection may not always be device-related.</p>																																					
	<table border="1"> <thead> <tr> <th rowspan="4">Patient Residual Harm Category</th> <th colspan="2">Quantification of Residual Risks</th> </tr> <tr> <th>PMS Complaints (01 January 2016 – 31 March 2025)</th> <th>Post Market Clinical Follow-Up Activity Events</th> </tr> <tr> <th>Units Sold: 112,258</th> <th>Units Studied: 107</th> </tr> <tr> <th># of Cases Per Event</th> <th># of Cases Per Event</th> </tr> </thead> <tbody> <tr> <td>Allergic Reaction</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> <tr> <td>Bleeding</td> <td>1 Event in 16,000 Cases.</td> <td>Not Reported.</td> </tr> <tr> <td>Cardiac Event</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> <tr> <td>Embolism</td> <td>1 Event in 100,000 Cases.</td> <td>Not Reported.</td> </tr> <tr> <td>Infection</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> <tr> <td>Perforation</td> <td>1 Event in 100,000 Cases.</td> <td>Not Reported.</td> </tr> <tr> <td>Stenosis</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> <tr> <td>Tissue Injury</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> <tr> <td>Thrombosis</td> <td>Not Reported.</td> <td>Not Reported.</td> </tr> </tbody> </table>		Patient Residual Harm Category	Quantification of Residual Risks		PMS Complaints (01 January 2016 – 31 March 2025)	Post Market Clinical Follow-Up Activity Events	Units Sold: 112,258	Units Studied: 107	# of Cases Per Event	# of Cases Per Event	Allergic Reaction	Not Reported.	Not Reported.	Bleeding	1 Event in 16,000 Cases.	Not Reported.	Cardiac Event	Not Reported.	Not Reported.	Embolism	1 Event in 100,000 Cases.	Not Reported.	Infection	Not Reported.	Not Reported.	Perforation	1 Event in 100,000 Cases.	Not Reported.	Stenosis	Not Reported.	Not Reported.	Tissue Injury	Not Reported.	Not Reported.	Thrombosis	Not Reported.	Not Reported.
	Patient Residual Harm Category	Quantification of Residual Risks																																				
		PMS Complaints (01 January 2016 – 31 March 2025)		Post Market Clinical Follow-Up Activity Events																																		
		Units Sold: 112,258		Units Studied: 107																																		
		# of Cases Per Event	# of Cases Per Event																																			
	Allergic Reaction	Not Reported.	Not Reported.																																			
	Bleeding	1 Event in 16,000 Cases.	Not Reported.																																			
Cardiac Event	Not Reported.	Not Reported.																																				
Embolism	1 Event in 100,000 Cases.	Not Reported.																																				
Infection	Not Reported.	Not Reported.																																				
Perforation	1 Event in 100,000 Cases.	Not Reported.																																				
Stenosis	Not Reported.	Not Reported.																																				
Tissue Injury	Not Reported.	Not Reported.																																				
Thrombosis	Not Reported.	Not Reported.																																				
<p>Warnings and precautions</p>	<p>The below are warnings, precautions, or measures to be taken by patient:</p>																																					
	<ul style="list-style-type: none"> <li>• To reduce the risk of bacteria entering the catheter, wear a mask over your nose and mouth whenever the catheter is accessed.</li> <li>• Keep the catheter dressing clean and dry. The dressing should be changed by a medical professional at each dialysis session.</li> </ul>																																					

	<ul style="list-style-type: none"> <li>• Avoid letting the catheter or catheter site go under water. Moisture near the catheter site can potentially lead to an infection.</li> <li>• Ask the doctor to explain the signs and symptoms of catheter infection.</li> <li>• Never remove the cap at the end of the catheter. The cap and clamps of the catheter must be kept closed when not being used for dialysis.</li> </ul>
Summary of any field safety correction action (FSCA)	There were no recalls for the device between 01 April 2024 to 31 March 2025.

## 5. Summary of clinical evaluation and post-market clinical follow-up

<b>Clinical background of the device</b>
<p>The Split Stream® catheter has been available since 2003. The CE Mark was received in 2003. US FDA clearance was in February 2003. All models included are planned for distribution in the European Union.</p>
<b>Clinical evidence for CE-marking</b>
<p>The clinical literature review found 3 articles relating to the safety and/or performance of the subject device when used as intended. These articles included approximately 179 cases. Three patient level data activities received information on 107 catheters.</p> <p>Findings from the clinical literature and clinical survey support the performance of the subject device. All data on the Split Stream® catheter has been evaluated. The benefits of the subject device outweigh the risks when the device is used as intended. The benefit of the device is allowing hemodialysis in patients in whom other therapies or conservative care are not desirable by the physician.</p>
<b>Safety</b>
<p>There is sufficient data to prove conformity to the applicable requirements. The device is safe and performs as intended and claimed by Medcomp. The device is state of the art for allowing long-term vascular access for hemodialysis in adult patients.</p> <p>Medcomp has reviewed:</p> <ul style="list-style-type: none"> <li>• Post-Market Data</li> <li>• Medcomp Information Materials</li> <li>• Risk Management Documentation</li> </ul> <p>The risks are appropriately displayed and consistent with the state of the art. The risks associated with the device are acceptable when weighed against the benefits. There were 285 complaints for 61,824 units sold from 01 January 2020 to 31 March 2025. The complaint rate is 0.46%.</p>

## 6. Possible therapeutic alternatives

When considering alternative treatments, it is recommended to contact your healthcare professional who can consider your individual situation. The Kidney Disease Outcomes Quality Initiative (KDOQI) 2019 clinical practice guidelines have been used to support the below recommendations for treatments.

Therapy	Benefits	Disadvantages	Key Risks
AV Fistula	<ul style="list-style-type: none"> <li>• Permanent solution.</li> <li>• Lower complication rate than catheter.</li> </ul>	<ul style="list-style-type: none"> <li>• Requires time.</li> <li>• Patients must sometimes self-needle stick.</li> </ul>	<ul style="list-style-type: none"> <li>• Stenosis</li> <li>• Thrombosis</li> <li>• Aneurysm</li> <li>• Pulmonary hypertension</li> <li>• Steal Syndrome</li> <li>• Septicemia</li> </ul>
Hemodialysis Catheter	<ul style="list-style-type: none"> <li>• Useful for quick access.</li> <li>• Can be used as a bridge between therapies.</li> </ul>	<ul style="list-style-type: none"> <li>• Not permanent.</li> <li>• Catheter dysfunction can happen.</li> <li>• Benefit may not be the same for everyone.</li> </ul>	<ul style="list-style-type: none"> <li>• Post-procedural bleeding</li> <li>• Infection</li> <li>• Thrombosis</li> <li>• Decreased blood flow in dysfunctional catheter</li> <li>• Cardiovascular events</li> <li>• Fibrin sheath formation around catheter</li> <li>• Septicemia</li> </ul>
Peritoneal Dialysis	<ul style="list-style-type: none"> <li>• Less restrictive diet than hemodialysis.</li> <li>• Does not require hospitalization.</li> </ul>	<ul style="list-style-type: none"> <li>• Clearance of impurities is limited by flow and space.</li> </ul>	<ul style="list-style-type: none"> <li>• Peritonitis</li> <li>• Septicemia</li> <li>• Fluid overload</li> </ul>
Kidney Transplant	<ul style="list-style-type: none"> <li>• Better quality of life.</li> <li>• Lower risk of death.</li> <li>• Fewer dietary restrictions.</li> </ul>	<ul style="list-style-type: none"> <li>• Requires a donor.</li> <li>• More risky for certain groups.</li> <li>• Patient must take medication for life.</li> <li>• Medication has side effects.</li> </ul>	<ul style="list-style-type: none"> <li>• Thrombosis</li> <li>• Hemorrhage</li> <li>• Ureteral blockage</li> <li>• Infection</li> <li>• Organ rejection</li> <li>• Death</li> <li>• Myocardial infarction</li> <li>• Stroke</li> </ul>
Comprehensive Conservative Care	<ul style="list-style-type: none"> <li>• Less imposed symptom burden.</li> <li>• Preserves life satisfaction.</li> </ul>	<ul style="list-style-type: none"> <li>• May aggravate clinical condition.</li> <li>• Not designed to treat.</li> </ul>	<ul style="list-style-type: none"> <li>• Treatment may not actually minimize risks associated with CKD.</li> </ul>

## 7. Suggested training for users

The catheter should be inserted, manipulated, and removed by a qualified, licensed physician or other qualified health care professional under the direction of a physician. In certain

circumstances, patients who may be suitable for home hemodialysis may manipulate the external connections of the catheter.

Consult International Society of Hemodialysis guidelines. If home dialysis is recommended, you will undergo thorough training. The objectives of the training program are:

- 1) Give you information to dialyze safely at home.
- 2) Enable you to monitor and manage your disease.
- 3) Help you cope with fears and restrictions of home hemodialysis.

The ideal nurse trainer-to-patient ratio is typically 1:1. A training schedule will be created. Training will be individualized to your needs.

Abbreviation	Definition
AV	Arteriovenous
CE	Conformité Européenne (European Conformity)
CKD	Chronic Kidney Disease
cm	centimeter
CMR	Carcinogenic, mutagenic, reprotoxic
dba	Doing Business As
F	French (thickness of catheter)
FDA	Food and Drug Administration
FSCA	Field Safety Corrective Action
KDOQI	Kidney Disease Outcomes Quality Initiative
PA	Pennsylvania
SSCP	Summary of Safety and Clinical Performance
USA	United States of America
w/w	Weight over Weight

Add copy to 'MDR Documentation' (Initial & Date) DM 25SEP2025